



DIESSSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME HELICOBACTER PYLORI Ag - CHORUS

Code	81063		
Lot No.	026	2025-04	2024-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D6960	Helicobacter pylori Ag	026	2025-04	2X6
B6960	CALIBRATORE Helicobacter pylori Ag	026	2025-06	2
C6960	CONTROL SERUM Helicobacter pylori Ag	026	2025-06	3
G6960	H. pylori Ag Stool Diluent/Negative Control	D1061700	2025-05	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	250	≥ 100
Control + (Index)	4,3	2,8-5,2
Stool Diluent/Negative Control (Index)	0,2	<0,5

CALIBRATOR

9	3	2	0	2	6	6	0	0	0	0	0	2	5	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature HIB

93202660000250000
Helicobacter pylori Ag
CALIBRATOR 0.250 mL [I]VD
LOT 026
 2025-06 2°C-8°C
DIESSSE Monteriggioni Italy CE

CONTROL SERUM

9	3	2	0	2	6	6	2	8	5	2	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature HIB

932026628520000001
Helicobacter pylori Ag
CONTROL+ 0.250 mL [I]VD
LOT 026
 2025-06 2°C-8°C
DIESSSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. HIB

DATE 22/04/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.