

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME TETANUS IgG - CHORUS

Code 81270

Lot No. 024



2025-04



2024-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D9660	Tetanus IgG	024	2025-04	6X6
B9660	CALIBRATORE Tetanus IgG	024	2025-04	1
C9660	CONTROL SERUM Tetanus IgG	024	2025-04	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	427	≥ 100
Control (mIU/mL)	705,8	475,0-1136,4

CALIBRATOR

9	0	2	0	2	4	6	2	2	0	3	4	3	7	4	5	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature




CONTROL SERUM

9	0	2	0	2	4	6	1	9	3	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.

DATE 29/04/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.