

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME CHLAMYDOPHILA PNEUMONIAE IgM - CHORUS

Code	81251
Lot No.	024
	2025-04
	2024-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D8600	Chlamydomphila pneumoniae IgM	024	2025-04	6X6
B8600	CALIBRATORE Chlamydomphila p. IgM	024	2025-04	1
C8600	CONTROL SERUM Chlamydomphila p. IgM	024	2025-04	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	623	≥ 100
Control Serum (Index)	1,6	1,3-2,5
Negative Control (Index)	0,2	≤ 0.7

CALIBRATOR

2	3	0	0	2	4	6	0	0	0	0	0	5	3	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature FB



CONTROL SERUM

2	3	0	0	2	4	6	1	3	2	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature FB



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents [Signature]

DATE 06/05/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.