

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CARDIOLIPIN-M - CHORUS

Code	86048
Lot No.	024
	2025-04
	2024-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D3331	CARDIOLIPIN-M	024	2025-04	6X6
B3331	CALIBRATORE CARDIOLIPIN-M	024	2025-04	1
C3331	CONTROL SERUM CARDIOLIPIN-M	024	2025-04	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	550	≥ 100
Control Serum (MPLU/mL)	51,3	38,2-76,8

CALIBRATOR

3	3	6	0	2	4	6	4	1	6	2	7	3	7	4	6	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *FB*



CONTROL SERUM

3	3	6	0	2	4	6	1	9	3	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *FB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *Blue*

DATE 02/05/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.