

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554



QUALITY CONTROL CERTIFICATE

KIT NAME ENA-6 S - CHORUS

Code	86012		
Lot No.	004	2025-01	2024-01

REAGENTS

Label	Description	Lot.	Exp.	NO
D3132	ENA-6 S	004	2025-01	6X6
B3132	CALIBRATORE ENA-6 S	004	2025-01	1
C3132	CONTROL SERUM ENA-S 6	004	2025-01	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	438	≥ 100
Control Serum (ndx)	2,0	1,5-2,8

CALIBRATOR

1	0	2	0	0	4	6	0	0	0	0	0	0	4	6	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES Operator Signature *[Signature]*



CONTROL SERUM

1	0	2	0	0	4	6	1	5	2	8	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *[Signature]*



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. *[Signature]*

DATE 24.04.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.