



DIESSSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME SS-A - CHORUS

Code	86018		
Lot No.	025	2025-07	2024-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D3181	SS-A	025	2025-07	6X6
B3181	CALIBRATORE SS-A	025	2025-07	1
C3181	CONTROL SERUM SS-A	025	2025-07	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	453	≥ 100
Control Serum (AU/mL)	49,4	30,0-64,4

CALIBRATOR

1	9	3	0	2	5	7	1	2	0	3	5	3	5	5	3	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *MB*

193025712035355300
SS-A
CALIBRATOR 0.175 mL [IVD]
LOT 025
 2025-07 2°C-8°C
DIESSSE Monteriggioni Italy CE

CONTROL SERUM

1	9	3	0	2	5	7	2	0	3	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *MB*

193025720370000001
SS-A
CONTROL+ 0.425 mL [IVD]
LOT 025
 2025-07 2°C-8°C
DIESSSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*
DATE *29.06.2024*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!
This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.