



DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME BORDETELLA PERTUSSIS TOXIN IgG - CHORUS

Code	81272		
Lot No.	025	2025-04	2024-04

REAGENTS

Label	Description	Lot.	Exp	No
D2100	BORDETELLA PERTUSSIS TOXIN IgG	025	2025-04	6x6
B2100	CALIBRATORE BORDETELLA P.TOXIN IgG	025	2025-08	1
C2100	CONTROL SERUM BORDETELLA P.TOXIN IgG	025	2025-08	1

CONTROLS

	Found results	Expected results
CALIBRATOR (mOD)	457	≥100
CONTROL SERUM (IU/ml)	175,4	114,3-260,0

CALIBRATOR

9	6	4	0	2	5	6	1	1	8	3	2	3	6	4	2	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *Olivero*

964025611832364200  
Bordetella pertussis  
Toxin IgG  
CALIBRATOR 0.450 mL IVD  
LOT 025  
2025-08 2°C-8°C  
DIESSE Monteriggioni Italy CE

CONTROL SERUM

9	6	4	0	2	5	6	2	0	3	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *Olivero*

964025620360000001  
Bordetella pertussis  
Toxin IgG  
CONTROL+ 0.500 mL IVD  
LOT 025  
2025-08 2°C-8°C  
DIESSE Monteriggioni Italy CE

**This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.**

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *Olivero*

DATE 30-06-2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.