

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

DIESSE

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME TICK-BORNE ENCEPHALITIS VIRUS IgM - CHORUS

Code 81277

Lot No. 024



2025-04



2024-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D0160	TICK-BORNE ENCEPHALITIS VIRUS IgM	024	2025-04	6X6
B0160	CALIBRATORE TICK-BORNE ENCEPHALITIS VIRUS IgM	024	2025-04	1
C0160	CONTROL SERUM TICK-BORNE ENCEPHALITIS VIRUS IgM	024	2025-04	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	370	≥ 100
Control (index)	2,8	1,7-3,1

CALIBRATOR


9 3 7 0 2 4 6 0 0 0 0 0 4 2 0 0 0 0

Label checked with barcode reader

YES



Operator Signature *FB*



CONTROL SERUM

9 3 7 0 2 4 6 1 7 3 1 0 0 0 0 0 0 1

Label checked with barcode reader

YES



Operator Signature *FB*



This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. *FB*

DATE 30/04/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.