



DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME Epstein-Barr VCA IgG - CHORUS

Code	81055		
Lot No.	018	2025-06	2024-03

REAGENTS

Label	Description	Lot	Exp.	NO
D2341	Epstein-Barr VCA IgG	018	2025-06	6X6
B2342	CALIBRATORE Epstein-Barr VCA IgG	018	2025-06	1
C2341	CONTROL SERUM Epstein-Barr VCA IgG	018	2025-06	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	320	≥ 100
Control Serum (INDEX)	3,6	2,7-5,0
Negative Control (INDEX)	0,0	<0,7

CALIBRATOR

0	0	5	0	1	8	7	0	0	0	0	0	2	6	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *Osteo*



CONTROL SERUM

0	0	5	0	1	8	7	2	7	5	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *Osteo*



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *Osteo*

DATE *07.05.2024*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!
This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.