



QUALITY CONTROL CERTIFICATE

KIT NAME MYCOPLASMA PNEUMONIAE IgM - CHORUS

Code 81035

Lot No. 012



2025-05



2024-02

REAGENTS

Label	Description	Lot.	Exp.	NO
D2262	MYCOPLASMA PNEUMONIAE IgM	012	2025-05	6x6
B2262	CALIBRATORE MYCOPLASMA PNEUMONIAE IgM	012	2025-06	1
C2262	CONTROL SERUM MYCOPLASMA PNEUMONIAE IgM	012	2025-06	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	420	≥ 100
Control Serum (INDEX)	1,5	1,1-2,0
Negative Control (INDEX)	0,1	$\leq 0,7$

CALIBRATOR

0	6	7	0	1	2	7	0	0	0	0	0	0	6	6	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature

Operator Signature



CONTROL SERUM

0	6	7	0	1	2	7	1	1	2	0	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature

Operator Signature



**This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.**

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

Signature

DATE 10-05-2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.