

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME PSA

Code 86550
Lot No. 034 2025-05 2024-05

REAGENTS

Label	Description	Lot.	Exp.	NO
D2020	PSA	034	2025-05	6X6
B2020	CALIBRATOR PSA	034	2026-06	1
C2020	CONTROL SERUM PSA	034	2026-06	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	360	≥ 100
Control Serum (ng/ml)	16,5	13,7-19,1

CALIBRATOR

9 4 8 0 3 4 6 1 4 9 5 7 3 0 6 4 0 0

Label checked with barcode reader YES Operator Signature *MB*

948034614957306400
PSA
CALIBRATOR 0.280 mL IVD
LOT 034
2026-06 2°C-8°C
DIESSE Monteriggioni Italy CE 0123

CONTROL SERUM

9 4 8 0 3 4 6 2 8 3 8 0 0 0 0 0 0 1

Label checked with barcode reader YES Operator Signature *MB*

948034628380000001
PSA
CONTROL 0.800 mL IVD
LOT 034
2026-06 2°C-8°C
DIESSE Monteriggioni Italy CE 0123

This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *MB*

DATE 16.05.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!
This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.