



QUALITY CONTROL CERTIFICATE

KIT NAME tTg-G - CHORUS

Code	86060		
Lot No.	019	2025-03	2024-03

REAGENTS

Label	Description	Lot.	Exp.	NO
D345I	tTg-G	019	2025-03	6X6
B345I	CALIBRATORE tTg-G	019	2025-03	1
C345I	CONTROL SERUM tTg-G	019	2025-03	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	416	≥ 100
Control Serum (AU/mL)	47,4	39,0-92,0

CALIBRATOR

3	4	0	0	1	9	6	1	1	4	3	4	3	5	3	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature



340019611434353900
tTg-G

CALIBRATOR 0.175 mL [IVD]

LOT 019


2025-03 2°C-8°C

DIESSE Monteriggioni Italy

CONTROL SERUM

3	4	0	0	1	9	6	2	0	3	8	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature



340019620380000001
tTg-G

CONTROL 0.425 mL [IVD]

LOT 019

2025-03 2°C-8°C

DIESSE Monteriggioni Italy

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 1605-2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.