

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CALPROTECTIN 3RD GENERATION- CHORUS

Code 86603

Lot No. 018



2025-03



2024-03

REAGENTS

Label	Description	Lot.	Exp.	NO
D8430	CALPROTECTIN 3 RD GENERATION	018	2025-03	6X6
B8430	CALIBRATORE CALPROTECTIN 3 RD GENERATION	018	2025-03	1
C8430	CONTROL CALPROTECTIN 3 RD GENERATION	018	2025-03	1
G8430	SAMPLE DILUENT/NEGATIVE CONTROL CALPROTECTIN 3 RD GENERATION	018	2025-03	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	388	≥ 100
Control (mg/Kg)	434,8	293,2-850,0
sample dil/neg control Calp 3 RD (mg/Kg)	33,8	< 51,0

CALIBRATOR

9	9	3	0	1	8	6	3	2	4	3	6	3	3	4	4	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature MB



CONTROL SERUM

9	9	3	0	1	8	6	2	1	4	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature MB



This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED BY RESPONSIBLE HEAD, Quality Control Reagents

DATE 16.05.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.