

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

DIESSE

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME TICK-BORNE ENCEPHALITIS VIRUS IgG - CHORUS

Code 81276

Lot No. 034



2025-05



2024-05

REAGENTS

Label	Description	Lot.	Exp.	NO
D0140	TICK-BORNE ENCEPHALITIS VIRUS IgG	034	2025-05	6X6
B0141	CALIBRATORE TICK-BORNE ENCEPHALITIS VIRUS IgG	034	2025-05	1
C0141	CONTROL SERUM TICK-BORNE ENCEPHALITIS VIRUS IgG	034	2025-05	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	390	≥ 100
Control (AU/mL)	47,0	41,3-60,6

CALIBRATOR


9	6	6	0	3	4	6	1	1	8	4	9	3	8	8	8	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *[Signature]*



CONTROL SERUM


9	6	6	0	3	4	6	1	9	3	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *[Signature]*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE 17-05-2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.