

QUALITY CONTROL CERTIFICATE

KIT NAME TETANUS IgG - CHORUS

Code	81270/12		
Lot No.	025	2025-04	2024-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D9660	Tetanus IgG	025	2025-04	2X6
B9660	CALIBRATORE Tetanus IgG	025	2025-04	1
C9660	CONTROL SERUM Tetanus IgG	025	2025-04	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	437	≥ 100
Control (mIU/mL)	775,6	541,7-1230,8

CALIBRATOR

9	0	2	0	2	5	6	1	1	7	2	9	4	0	5	5	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES Operator Signature



CONTROL SERUM

9	0	2	0	2	5	6	1	8	3	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature



This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 17-05-2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.