

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

DIESSE

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME dsDNA-G - CHORUS

Code 86032

Lot No. 026



2025-04



2024-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D3281	dsDNA-G	026	2025-04	6X6
B3281	CALIBRATORE dsDNA-G	026	2025-04	1
C3281	CONTROL SERUM dsDNA-G	026	2025-04	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	726	≥ 100
Control Serum (IU/mL)	54,6	34,1-73,1

CALIBRATOR


2	5	1	0	2	6	6	5	2	1	3	4	4	9	3	7	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *MB*



CONTROL SERUM


2	5	1	0	2	6	6	1	4	2	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *MB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE

20-05-2024

[Handwritten signature]

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.