

QUALITY CONTROL CERTIFICATE

KIT NAME BORDETELLA PERTUSSIS TOXIN IgA - CHORUS

Code	81274		
Lot No.	034	2025-05	2024-05

REAGENTS

Label	Description	Lot.	Exp	No
D3440	BORDETELLA PERTUSSIS TOXIN IgA	034	2025-05	6X6
B3440	CALIBRATORE BORDETELLA P.TOXIN IgA	034	2025-05	1
C3440	CONTROL SERUM BORDETELLA P.TOXIN IgA	034	2025-05	1

CONTROLS

	Found results	Expected results
CALIBRATOR (mOD)	520	>100
CONTROL SERUM ( IU/ml)	26,0	17,5-30,0

CALIBRATOR

9	6	5	0	3	4	6	1	2	2	3	7	6	0	5	2	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES

Operator Signature

965034612237605200  
Bordetella pertussis  
Toxin IgA  
CALIBRATOR 0.450 mL IVD  
LOT 034  
2025-05  
DIESSE Monteriggioni Italy  
2°C-8°C  
CE  
B3440

CONTROL SERUM

9	6	5	0	3	4	6	1	2	2	2	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES

Operator Signature

965034612220000001  
Bordetella pertussis  
Toxin IgA  
CONTROL 0.900 mL IVD  
LOT 034  
2025-05  
DIESSE Monteriggioni Italy  
2°C-8°C  
CE  
C3440

*This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 21/05/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.