



## EU DECLARATION OF CONFORMITY

MANUFACTURER:	DIESSSE DIAGNOSTICA SENESE SPA STRADA DEI LAGHI 39 53035 MONTERIGGIONI (SI), ITALY
SINGLE REGISTRATION NUMBER	IT-MF-000013311
AUTHORIZED REPRESENTATIVE:	//
PRODUCT: CODE:	<b>CHORUS ELASTASE COPROCOLLECT 86116</b>
INTENDED PURPOSE:	Single-use device for manual collection of the amount of faeces required to perform the CHORUS CLIA FECAL ELASTASE (REF 82304) and CHORUS FECAL ELASTASE (REF 86114) kits. The device should only be used by professional laboratory personnel.
BASIC UDI-DI	803389132COPROEL-C00XT
UDI-DI	08033891328481
RISK CLASS:	CLASS A
CLASSIFICATION RULE:	RULE 5C
CONFORMITY ASSESSMENT ROUTE:	ARTICLE 17, ANNEX II and ANNEX III
WE HEREWITH DECLARE THAT THE ABOVE MENTIONED PRODUCT IS IN CONFORMITY WITH <b>REGULATION (EU) 2017/746</b> OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL ON IN VITRO DIAGNOSTIC MEDICAL DEVICE. ALL SUPPORTING DOCUMENTATION IS RETAINED UNDER THE PREMISES OF THE MANUFACTURER.	
REFERENCE TO ANY CS APPLIED:	NOT APPLICABLE
NOTIFIED BODY:	NOT NECESSARY
(EU) CERTIFICATE:	//
REVISION:	0
PLACE, DATE OF ISSUE:	MONTERIGGIONI, 23 MAY 2024
EXPIRY DATE:	//

**THE PRESENT EU DECLARATION OF CONFORMITY IS ISSUED UNDER THE SOLE AND EXCLUSIVE RESPONSIBILITY OF THE MANUFACTURER.**

*Maria Claudia Alcaro*

SIGNATURE:

MARIA CLAUDIA ALCARO  
PERSON RESPONSIBLE FOR THE REGULATORY  
COMPLIANCE

THIS IS THE CERTIFIED COPY OF THE ORIGINAL DOCUMENT STORED IN ARCHIVE OF DIESSE DIAGNOSTICA SENESE SPA

ISSUED: MONTERIGGIONI, 2024-05-23

*M. Stoczko*

MAGDALENA STOCZKO  
REGULATORY SUPERVISOR