

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554



QUALITY CONTROL CERTIFICATE

KIT NAME MYCOPLASMA PNEUMONIAE IgG - CHORUS

Code 81034

Lot No. 026

2025-04

2024-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D2252	MYCOPLASMA PNEUMONIAE IgG	026	2025-04	6x6
B2253	CALIBRATORE MYCOPLASMA PNEUMONIAE IgG	026	2025-04	1
C2252	CONTROL SERUM MYCOPLASMA PNEUMONIAE IgG	026	2025-04	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	320	≥ 100
Control Serum (UA/mL)	42.6	26.0-51.6
Negative Control (UA/mL)	<10.0	<10

CALIBRATOR

0	6	6	0	2	6	6	7	1	3	2	3	4	5	5	4	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature

[Signature]



CONTROL SERUM

0	6	6	0	2	6	6	1	6	2	9	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature

[Signature]



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

[Signature]

DATE 23/05/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive