

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CHLAMYDOPHILA PNEUMONIAE IgG - CHORUS

Code 81250

Lot No. 026



2025-04



2024-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D8590	Chlamydomphila pneumoniae IgG	026	2025-04	6X6
B8590	CALIBRATORE Chlamydomphila p. IgG	026	2025-04	1
C8590	CONTROL SERUM Chlamydomphila p. IgG	026	2025-04	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	470	≥ 100
Control Serum (AU/mL)	43,8	27,3-56,8
Negative Control (AU/mL)	$<10,0$	$\leq 10,0$

CALIBRATOR

2	2	2	0	2	6	6	5	1	2	2	3	4	3	4	8	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature

Open



CONTROL SERUM

2	2	2	0	2	6	6	1	6	3	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature

Open



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

Open

DATE 22-05-2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.