



DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME West Nile IgG -CHORUS

Code	81174		
Lot No.	033	2025-05	2024-05

REAGENTS

Label	Description	Lot.	Exp	No
D4080	West Nile IgG	033	2025-05	6X6
B4080	Calibratore West Nile IgG	033	2025-05	1
C4080	Control West Nile IgG	033	2025-05	1

CONTROLS

Found results	Expected results
---------------	------------------

Calibrator (mOD)	400	≥ 100
Control (Index)	2,3	1,6-3,0

CALIBRATOR

9	8	8	0	3	3	6	0	0	0	0	0	4	3	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature

98803360000430000  
West Nile Virus IgG  
CALIBRATOR 0.175 mL IVD  
LOT 033  
2025-06 2°C-8°C  
DIESSE Monteriggioni Italy CE  
B4080

CONTROL SERUM

9	8	8	0	3	3	6	1	6	3	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature

988033616300000001  
West Nile Virus IgG  
CONTROL+ 0.425 mL IVD  
LOT 033  
2025-06 2°C-8°C  
DIESSE Monteriggioni Italy CE  
C4080

**This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.**

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 22-05-2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.