



DIESSSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME MUMPS IgM CHORUS

Code	81076/12		
Lot No.	025	2025-04	2024-05

REAGENTS

Label	Description	Lot.	Exp.	NO
D2461	MUMPS IgM	025	2025-04	2x6
B2461	CALIBRATORE MUMPS IgM	025	2025-05	1
C2461	CONTROL SERUM MUMPS IgM	025	2025-05	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	480	≥ 100
Control Serum (Index)	2,0	1,5-2,7
Negative Control (Index)	0,0	< 0,7

CALIBRATOR

1	3	2	0	2	5	6	0	0	0	0	0	4	8	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES Operator Signature *Peter*



CONTROL SERUM

1	3	2	0	2	5	6	1	5	2	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *aw*



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *Florian*

DATE *26.05.2024*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!
This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.