

QUALITY CONTROL CERTIFICATE

KIT NAME MUMPS IgM CHORUS

Code	81076
Lot No.	025
	2025-04
	2024-05

REAGENTS

Label	Description	Lot.	Exp.	NO
D2461	MUMPS IgM	025	2025-04	6x6
B2461	CALIBRATORE MUMPS IgM	025	2025-05	1
C2461	CONTROL SERUM MUMPS IgM	025	2025-05	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	480	≥ 100
Control Serum (Index)	2,0	1,5-2,7
Negative Control (Index)	0,0	< 0,7

CALIBRATOR

1	3	2	0	2	5	6	0	0	0	0	0	4	8	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *[Signature]*

132025600000480000  
Mumps IgM  
CALIBRATOR 0.176 mL IVD  
LOT 025  
 2025-05  
DIESSA Monteriggioni Italy CE  
B2461

CONTROL SERUM

1	3	2	0	2	5	6	1	5	2	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *[Signature]*

132025615270000001  
Mumps IgM  
CONTROL 0.425 mL IVD  
LOT 025  
 2025-05  
DIESSA Monteriggioni Italy CE  
C2461

*This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE 24/05/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.