



DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME tTg-G - CHORUS

Code	86060		
Lot No.	020	2025-03	2024-03

REAGENTS

Label	Description	Lot.	Exp.	NO
D345I	tTg-G	020	2025-03	6X6
B345I	CALIBRATORE tTg-G	020	2025-03	1
C345I	CONTROL SERUM tTg-G	020	2025-03	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	420	≥ 100
Control Serum (AU/mL)	60,3	49,1-100,0

CALIBRATOR

3	4	0	0	2	0	6	1	1	6	2	7	3	0	4	4	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *Phen*

340020611627304400
tTg-G
CALIBRATOR 0.178 mL IVD
LOT 020
2025-03 2°C-8°C
DIESSE Monteriggioni Italy CE

CONTROL SERUM

3	4	0	0	2	0	6	2	3	4	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *Phen*

340020623440000001
tTg-G
CONTROL 0.425 mL IVD
LOT 020
2025-03 2°C-8°C
DIESSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *Phen*

DATE 26.05.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!
This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.