

QUALITY CONTROL CERTIFICATE

KIT NAME POLIOVIRUS IgG - CHORUS

| | | | |
|---------|----------|---------|---------|
| Code | 81264/12 | | |
| Lot No. | 035 | 2025-05 | 2024-05 |

REAGENTS

| Label | Description | Lot. | Exp. | NO |
|-------|------------------------------|------|---------|-----|
| D9630 | Poliovirus IgG | 035 | 2025-05 | 2X6 |
| B9630 | CALIBRATORE Poliovirus IgG | 035 | 2025-05 | 1 |
| C9630 | CONTROL SERUM Poliovirus IgG | 035 | 2025-05 | 1 |


CONTROLS

| | Found results | Expected results |
|------------------|---------------|------------------|
| Calibrator (mOD) | 640 | ≥ 100 |
| Control (Index) | 1,3 | 1,1-2,0 |

CALIBRATOR

| | | | | | | | | | | | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 9 | 0 | 1 | 0 | 3 | 5 | 6 | 0 | 0 | 0 | 0 | 0 | 6 | 5 | 0 | 0 | 0 | 0 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|


Label checked with barcode reader YES Operator Signature SB



CONTROL SERUM

| | | | | | | | | | | | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 9 | 0 | 1 | 0 | 3 | 5 | 6 | 1 | 1 | 2 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 1 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader YES Operator Signature SB



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... SB

DATE 24.05.2025

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.