

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

### QUALITY CONTROL CERTIFICATE

KIT NAME CHORUS PROMONITOR ANTI-INFLIXIMAB

Code	86701		
Lot No.	033	2025-05	2024-05

### REAGENTS

Label	Description	Lot.	Exp	No
D4320	PROMONITOR ANTI-INFLIXIMAB	033	2025-05	6X6
B4320	CALIBRATOR ANTI-INFLIXIMAB	033	2025-06	1
C4320	CONTROL ANTI-INFLIXIMAB	033	2025-06	1

### CONTROLS

Found results	Expected results
---------------	------------------

Calibrator (mOD)	430	$\geq 100$
Control (AU/ml)	45,4	33,8-53,7

### CALIBRATOR


2	7	5	0	3	3	6	5	2	4	4	4	2	6	7	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature MB



### CONTROL SERUM


2	7	5	0	3	3	6	3	1	4	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature MB



This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 29/05/2024

### WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.