



DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME MEASLES IgM CHORUS

Code 81073

Lot No. 017 2025-06 2024-05

REAGENTS

Label	Description	Lot.	Exp.	NO
D2442	MEASLES IgM	017	2025-06	6x6
B2442	CALIBRATORE MEASLES IgM	017	2025-06	2
C2442	CONTROL SERUM MEASLES IgM	017	2025-06	2

CONTROLS

	Found results	Expected results
Calibrator (mOD)	597	≥ 100
Control Serum (Index)	2,0	1,4-2,6
Negative Control (Index)	0,0	< 0,7

CALIBRATOR

1 3 3 0 1 7 7 0 0 0 0 5 0 0 0 0 0

Label checked with barcode reader YES  Operator Signature *Ben*

13301770000500000  
Measles IgM  
CALIBRATOR 0.175 mL [IVD]  
LOT 017  
2025-06 2°C-8°C  
DIESSE Monteriggioni Italy CE

CONTROL SERUM

1 3 3 0 1 7 7 1 4 2 6 0 0 0 0 0 0 1

Label checked with barcode reader YES  Operator Signature *Ben*

13301771426000001  
Measles IgM  
CONTROL 0.425 mL [IVD]  
LOT 017  
2025-06 2°C-8°C  
DIESSE Monteriggioni Italy CE

*This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *Ben*

DATE *30.05.2024*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.