



QUALITY CONTROL CERTIFICATE

KIT NAME DENGUE IgG - CHORUS

Code	81154		
Lot No.	033	2025-05	2024-05

REAGENTS

Label	Description	Lot.	Exp.	NO
D5730	Dengue IgG	033	2025-05	6X6
B5730	Calibratore Dengue IgG	033	2025-05	1
C5730	Control Dengue IgG	033	2025-05	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	560	≥ 100
Control (Index)	2,4	1,8-3,3

CALIBRATOR

1	0	8	0	3	3	6	0	0	0	0	0	0	4	0	0	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *Alena*

10803360000400000
Dengue IgG
CALIBRATOR 0.175 mL IVD
LOT 033
2025-05
DIESSE Monteriggioni Italy CE

CONTROL SERUM

1	0	8	0	3	3	6	1	8	3	3	0	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *Alena*

108033618330000001
Dengue IgG
CONTROL+ 0.425 mL IVD
LOT 033
2025-05
DIESSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *Alena*

DATE 30.05.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.