



QUALITY CONTROL CERTIFICATE

KIT NAME 88000-F85 CELERY

Code	88000-F85			
Lot No.	035	2024-10	2024-05	

REAGENTS

Label	Description	Lot.	Exp.	NO
D0961	88000-F85 CELERY	035	2024-10	2X6
B0960	CALIBRATORE 88000-F85 CELERY	035	2025-06	1
C0960	CONTROL SERUM 88000-F85 CELERY	035	2025-06	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	300	≥ 100
Control Serum (KU/L)	10,7	6,6-15,0

CALIBRATOR

6	8	3	0	3	5	3	5	1	9	5	8	2	7	7	8	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature

683035351958277800
Celery
CALIBRATOR 0.250 mL IVD
LOT 035
2025-06 2°C 8°C
DIESSE Monteriggioni Italy CE

CONTROL SERUM

6	8	3	0	3	5	3	2	7	5	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature

683035327500000001
Celery
CONTROL 0.450 mL IVD
LOT 035
2025-06 2°C 8°C
DIESSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 31/05/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.