

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME MEASLES IgG CHORUS

Code 81072

Lot No. 026

2025-07

2024-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D2431	MEASLES IgG	026	2025-07	6X6
B2432	CALIBRATORE MEASLES IgG	026	2025-07	1
C2431	CONTROL SERUM MEASLES IgG	026	2025-07	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	440	≥ 100
Control Serum (Index)	2,5	1,8-3,3
Negative Control (Index)	0,0	$< 0,7$


CALIBRATOR

0 0 9 0 2 6 7 0 0 0 0 0 4 0 0 0 0 0

Label checked with barcode reader

YES ☒

Operator Signature *FB*



CONTROL SERUM

0 0 9 0 2 6 7 1 8 3 3 0 0 0 0 0 0 1

Label checked with barcode reader

YES ☒

Operator Signature *FB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. *[Signature]*

DATE 03/06/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.