

DIESE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME TETANUS IgG - CHORUS

Code 81270/12

Lot No. 026 2025-04 2024-04

REAGENTS

Label	Description	Lot	Exp.	NO
D9660	Tetanus IgG	026	2025-04	2X6
B9660	CALIBRATORE Tetanus IgG	026	2025-04	1
C9660	CONTROL SERUM Tetanus IgG	026	2025-04	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	384	≥ 100
Control (mIU/mL)	690,1	541,7-1214,3

CALIBRATOR

9 0 2 0 2 6 6 1 1 7 2 9 4 0 5 7 0 0

Label checked with barcode reader YES  Operator Signature

902026611729405700  
Tetanus IgG  
CALIBRATOR 0.175 mL IVD  
LOT 026  
2025-04 2°C-8°C  
DIESE Monteriggioni Italy CE

CONTROL SERUM

9 0 2 0 2 6 6 1 8 3 3 0 0 0 0 0 0 1

Label checked with barcode reader YES  Operator Signature

902026618330000001  
Tetanus IgG  
CONTROL 0.425 mL IVD  
LOT 026  
2025-04 2°C-8°C  
DIESE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 03.06.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.