

QUALITY CONTROL CERTIFICATE

KIT NAME TETANUS IgG - CHORUS

Code	81270		
Lot No.	026	2025-04	2024-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D9660	Tetanus IgG	026	2025-04	6X6
B9660	CALIBRATORE Tetanus IgG	026	2025-04	1
C9660	CONTROL SERUM Tetanus IgG	026	2025-04	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	384	≥ 100
Control (mIU/mL)	690,1	541,7-1214,3

CALIBRATOR

9	0	2	0	2	6	6	1	1	7	2	9	4	0	5	7	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature

902026611729405700
Tetanus IgG
CALIBRATOR 0.175 mL IVD
LOT 026
2025-04
DIESSSE Monteriggioni Italy CE

CONTROL SERUM

9	0	2	0	2	6	6	1	8	3	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature

902026618330000001
Tetanus IgG
CONTROL 0.425 mL IVD
LOT 026
2025-04
DIESSSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 03.06.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.