



DIESSSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME snRNP-C - CHORUS

Code	86092		
Lot No.	033	2025-05	2024-05

REAGENTS

Label	Description	Lot.	Exp.	NO
D7441	snRNP-C	033	2025-05	6X6
B7441	CALIBRATORE snRNP-C	033	2025-05	1
C7441	CONTROL SERUM snRNP-C	033	2025-05	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	380	≥ 100
Control Serum (AU/mL)	73,5	45,0-100,0

CALIBRATOR

1	9	1	0	3	3	6	1	1	6	2	6	3	3	4	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature MB

191033611626334000
snRNP-C
CALIBRATOR 0.175 mL IVD
LOT 033
2025-06 2°C
DIESSSE Monteriggioni Italy CE

CONTROL SERUM

1	9	1	0	3	3	6	2	1	4	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature MB

191033621400000001
snRNP-C
CONTROL 0.425 mL IVD
LOT 033
2025-06 2°C
DIESSSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents [Signature]

DATE 03/06/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.