



QUALITY CONTROL CERTIFICATE

KIT NAME Chorus Chlamydia trachomatis IgA

Code	81262			
Lot No.	033	2025-05	2024-05	

REAGENTS

Label	Description	Lot.	Exp.	NO
D9650	Chorus Chlamydia trachomatis IgA	033	2025-05	6X6
B9650	CALIBRATOR Chlamydia trachomatis IgA	033	2025-05	1
C9650	POSITIVE CONTROL Chlamydia trachomatis IgA	033	2025-05	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	290	≥ 100
Positive Control (Index)	3,4	2,5-4,7

CALIBRATOR

2	3	2	0	3	3	6	0	0	0	0	0	2	8	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES Operator Signature MB



CONTROL SERUM

2	3	2	0	3	3	6	2	5	4	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature MB



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents [Signature]

DATE 03/06/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!
This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.