



QUALITY CONTROL CERTIFICATE

KIT NAME **BORDETELLA PERTUSSIS TOXIN IgG - CHORUS**

Code	81272
Lot No.	034
	2025-05
	2024-05

REAGENTS

Label	Description	Lot.	Exp	No
D2100	BORDETELLA PERTUSSIS TOXIN IgG	034	2025-05	6x6
B2100	CALIBRATORE BORDETELLA P.TOXIN IgG	034	2025-09	1
C2100	CONTROL SERUM BORDETELLA P.TOXIN IgG	034	2025-09	1

CONTROLS

Found results	Expected results
---------------	------------------

CALIBRATOR (mOD)	400	≥100
CONTROL SERUM (IU/ml)	167,7	100,0-260,0

CALIBRATOR

9	6	4	0	3	4	6	1	1	9	3	2	3	7	4	2	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES

Operator Signature *Cohen*

964034611932374200  
Bordetella pertussis  
Toxin IgG  
CALIBRATOR 0.450 mL IVD  
LOT 034  
2025-09 2°C-8°C  
DIESSE Monteriggioni Italy CE

CONTROL SERUM

9	6	4	0	3	4	6	1	9	3	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES

Operator Signature *Cohen*

964034619360000001  
Bordetella pertussis  
Toxin IgG  
CONTROL 0.500 mL IVD  
LOT 034  
2025-09 2°C-8°C  
DIESSE Monteriggioni Italy CE

*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *Cohen*

DATE *06-06-2024*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.