

QUALITY CONTROL CERTIFICATE

KIT NAME LEGIONELLA PNEUMOPHILA 1-6 IgG -CHORUS

Code	81090		
Lot No.	033	2025-02	2024-05

REAGENTS

Label	Description	Lot	Exp.	NO
D7381	LEGIONELLA PNEUMOPHILA 1-6 IgG	033	2025-02	6X6
B7381	CALIBRATORE L PNEUMOPHILA 1-6 IgG	033	2025-02	1
C7381	CONTROL SERUM L PNEUMOPHILA 1-6 IgG	033	2025-02	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	400	≥ 100
Control Serum (INDEX)	3,4	2,2-4,1
Negative Control (INDEX)	0,0	<0,7

CALIBRATOR

0	6	3	0	3	3	5	0	0	0	0	0	0	3	2	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES Operator Signature



CONTROL SERUM

0	6	3	0	3	3	5	2	2	4	1	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature



This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 06.06.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.