

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554



QUALITY CONTROL CERTIFICATE

KIT NAME CARDIOLIPIN-M - CHORUS

Code 86048

Lot No. 033

2025-05

2024-05

REAGENTS

Label	Description	Lot	Exp.	NO
D3332	CARDIOLIPIN-M	033	2025-05	6X6
B3332	CALIBRATORE CARDIOLIPIN-M	033	2025-05	1
C3332	CONTROL SERUM CARDIOLIPIN-M	033	2025-05	1

CONTROLS

Found results	Expected results
---------------	------------------

Calibrator (mOD)	480	≥ 100
Control Serum (MPLU/mL)	55,6	32,5-80,0

CALIBRATOR


3	3	6	0	3	3	6	4	1	9	3	1	3	5	4	3	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *[Signature]*



CONTROL SERUM

3	3	6	0	3	3	6	2	0	3	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *[Signature]*



This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE 06-06-2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.