

QUALITY CONTROL CERTIFICATE

KIT NAME **Cenp-B - CHORUS**

Code	86028		
Lot No.	033	2025-05	2024-05

REAGENTS

Label	Description	Lot.	Exp.	NO
D3241	Cenp-B	033	2025-05	6X6
B3241	CALIBRATORE Cenp-B	033	2025-05	1
C3241	CONTROL SERUM Cenp-B	033	2025-05	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	510	≥ 100
Control Serum (AU/mL)	59,9	39,0-90,8

CALIBRATOR

1	9	6	0	3	3	6	1	1	7	2	7	3	5	4	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *FB*

196033611727354000
Cenp-B
CALIBRATOR 0.175 mL IVD
LOT 033
 2025-05 2°C 8°C
DIESSSE Monteriggioni Italy CE

CONTROL SERUM

1	9	6	0	3	3	6	2	0	3	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *FB*

196033620370000001
Cenp-B
CONTROL 0.425 mL IVD
LOT 033
 2025-05 2°C 8°C
DIESSSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *[Signature]*

DATE 07/06/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.