

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554



QUALITY CONTROL CERTIFICATE

KIT NAME CALPROTECTIN - CHORUS

Code	86600		
Lot No.	012-A	2025-02	2024-02

REAGENTS

Label	Description	Lot.	Exp.	NO
D9500	CALPROTECTIN	012	2025-02	2 X 6
B9500	CALIBRATORE CALPROTECTIN	012-1	2025-02	1
C9500	CONTROL CALPROTECTIN	012-1	2025-02	1
G9502	SAMPLE DILUENT	012-1	2026-02	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	360	≥ 100
Control (mg/kg)	165,7	94,4-250,0

CALIBRATOR

9	0	0	0	1	2	6	3	1	9	2	4	3	9	3	3	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *MB*

900012631924393300
Calprotectin
CALIBRATOR 0.300 mL TVB
LOT 012-1
2025-02 2°C
DIESSE Monteriggioni Italy CE

CONTROL SERUM

9	0	0	0	1	2	6	1	8	3	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *MB*

900012618330000001
Calprotectin
CONTROL 0.600 mL TVB
LOT 012-1
2025-02 2°C
DIESSE Monteriggioni Italy CE

*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*
DATE *07.06.2024*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!
This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.