

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME **BORRELIA IgM - CHORUS**

Code	81047		
Lot No.	042	2025-06	2024-06

REAGENTS

Label	Description	Lot	Exp.	NO
D2310	BORRELIA IgM	042	2025-06	6X6
B2310	CALIBRATORE BORRELIA IgM	042	2025-07	1
C2310	CONTROL BORRELIA IgM	042	2025-07	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	370	≥ 100
Control (Index)	3,5	2,3-4,3

CALIBRATOR

9	0	5	0	4	2	6	0	0	0	0	0	3	0	0	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES  Operator Signature *[Signature]*



CONTROL SERUM

9	0	5	0	4	2	6	2	3	4	3	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *[Signature]*



**This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.**

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE **07-06-2024**

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.