

DIESE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39_0577319554



QUALITY CONTROL CERTIFICATE

KIT NAME RUBELLA IgM - CHORUS

Code	81031		
Lot No.	025	2025-04	2024-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D2241	RUBELLA IgM	025	2025-04	6X6
B2241	CALIBRATORE RUBELLA IgM	025	2025-04	1
C2241	CONTROL SERUM	025	2025-04	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	450	≥ 100
Control Serum (INDEX)	2.6	1.8-3.3
Negative Control (INDEX)	0.0	< 0,7

CALIBRATOR

1	3	5	0	2	5	6	0	0	0	0	0	4	0	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *AB*

13502560000400000
Rubella IgM
CALIBRATOR 0.178 mL IVD
LOT 025
 2025-04 2°C-8°C
DIESE Monteriggioni Italy CE 0123 B2241

CONTROL SERUM

1	3	5	0	2	5	6	1	8	3	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *dh*

135025618330000001
Rubella IgM
CONTROL 0.425 mL IVD
LOT 025
 2025-04 2°C-8°C
DIESE Monteriggioni Italy CE 0123 C2241

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *Pleauer*

DATE *11/06/2024*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!
This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.

DIESSSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554



QUALITY CONTROL CERTIFICATE

KIT NAME RUBELLA IgM - CHORUS

Code 81031
Lot No. 025 2025-04 2024-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D2241	RUBELLA IgM	025	2025-04	6X6
B2241	CALIBRATORE RUBELLA IgM	025	2025-04	1
C2241	CONTROL SERUM	025	2025-04	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	450	≥ 100
Control Serum (INDEX)	2.6	1.8-3.3
Negative Control (INDEX)	0.0	< 0,7

CALIBRATOR

1 3 5 0 2 5 6 0 0 0 0 0 4 0 0 0 0 0

Label checked with barcode reader YES Operator Signature *AB*

13502560000400000
Rubella IgM
CALIBRATOR 0.176 mL [VD]
LOT 025
2025-04 2°C-8°C
DIESSSE Monteriggioni Italy CE 0123

CONTROL SERUM

1 3 5 0 2 5 6 1 8 3 3 0 0 0 0 0 0 0 1

Label checked with barcode reader YES Operator Signature *dh*

135025618330000001
Rubella IgM
CONTROL 0.426 mL [VD]
LOT 025
2025-04 2°C-8°C
DIESSSE Monteriggioni Italy CE 0123

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *Pleiner*

DATE *11/06/2024*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!
This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.