

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554



QUALITY CONTROL CERTIFICATE

KIT NAME HERPES SIMPLEX 1+2 IgM -CHORUS

| | | | |
|---------|-------|---------|---------|
| Code | 81021 | | |
| Lot No. | 016-A | 2025-03 | 2024-03 |

REAGENTS

| Label | Description | Lot. | Exp. | NO |
|-------|--------------------------------------|-------|---------|-----|
| D2222 | Herpes simplex 1+2 IgM | 016 | 2025-03 | 6x6 |
| B2222 | CALIBRATORE Herpes simplex 1+2 IgM | 016-1 | 2025-05 | 1 |
| C2222 | CONTROL SERUM Herpes simplex 1+2 IgM | 016-1 | 2025-05 | 1 |

CONTROLS

| | Found results | Expected results |
|--------------------------|---------------|------------------|
| Calibrator (mOD) | 445 | ≥ 100 |
| Control Serum (INDEX) | 1.7 | 1,4-2,7 |
| Negative Control (INDEX) | 0.0 | <0,7 |

CALIBRATOR

| | | | | | | | | | | | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 1 | 5 | 6 | 0 | 1 | 6 | 6 | 0 | 0 | 0 | 0 | 0 | 4 | 9 | 0 | 0 | 0 | 0 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader YES Operator Signature

15601660000490000
HERPES SIMPLEX 1+2 IgM
CALIBRATOR 0.175 mL IVD
LOT 016-1
 2025-05 2°C-8°C
DIESSE Monteriggioni Italy CE
B2222

CONTROL SERUM

| | | | | | | | | | | | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 1 | 5 | 6 | 0 | 1 | 6 | 6 | 1 | 4 | 2 | 7 | 0 | 0 | 0 | 0 | 0 | 0 | 1 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader YES Operator Signature

15601661427000001
HERPES SIMPLEX 1+2 IgM
CONTROL 0.425 mL IVD
LOT 016-1
 2025-05 2°C-8°C
DIESSE Monteriggioni Italy CE
C2222

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 11/06/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.