

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554



### QUALITY CONTROL CERTIFICATE

KIT NAME ANTI-CCP - CHORUS

Code 86094

Lot No. 023



2025-03



2024-03

### REAGENTS

| Label | Description            | Lot | Exp.    | NO  |
|-------|------------------------|-----|---------|-----|
| D8671 | ANTI-CCP               | 023 | 2025-03 | 6x6 |
| B8671 | CALIBRATORE ANTI-CCP   | 023 | 2025-03 | 1   |
| C8671 | CONTROL SERUM ANTI-CCP | 023 | 2025-03 | 1   |

### CONTROLS

|  | Found results | Expected results |
|--|---------------|------------------|
|--|---------------|------------------|

|                       |      |            |
|-----------------------|------|------------|
| Calibrator (mOD)      | 438  | $\geq 100$ |
| Control Serum (AU/mL) | 46,0 | 33,7-70,0  |


### CALIBRATOR

|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 2 | 5 | 6 | 0 | 2 | 3 | 6 | 3 | 2 | 5 | 4 | 5 | 3 | 9 | 6 | 1 | 0 | 0 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader

YES ☒

Operator Signature *MB*



### CONTROL SERUM

|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 2 | 5 | 6 | 0 | 2 | 3 | 6 | 1 | 8 | 3 | 3 | 0 | 0 | 0 | 0 | 0 | 0 | 1 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader

YES ☒

Operator Signature *MB*



This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *MB*

DATE 13.06.2024

### WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.