

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554



Pag 1/1

## QUALITY CONTROL CERTIFICATE

KIT NAME CARDIOLIPIN-M - CHORUS

Code 86048

Lot No. 034

2025-05

2024-05

## REAGENTS

Label	Description	Lot.	Exp.	NO
D3332	CARDIOLIPIN-M	034	2025-05	6X6
B3332	CALIBRATORE CARDIOLIPIN-M	034	2025-05	1
C3332	CONTROL SERUM CARDIOLIPIN-M	034	2025-05	1

## CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	470	$\geq 100$
Control Serum (MPLU/mL)	52,6	30,0-84,6

## CALIBRATOR


3	3	6	0	3	4	6	5	2	0	3	0	3	5	4	3	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



## CONTROL SERUM


3	3	6	0	3	4	6	2	0	3	8	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 13.06.2024

## WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.