

QUALITY CONTROL CERTIFICATE

KIT NAME CHLAMYDOPHILA PNEUMONIAE IgA - CHORUS

Code 81252

Lot No. 025

2025-04

2024-04

REAGENTS

Label	Description	Lot	Exp.	NO
D8610	Chlamydomphila pneumoniae IgA	025	2025-04	6X6
B8610	CALIBRATORE Chlamydomphila p. IgA	025	2025-04	1
C8610	CONTROL SERUM Chlamydomphila p. IgA	025	2025-04	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	547	≥ 100
Control Serum (AU/mL)	29,7	22,0-47,1
Negative Control (AU/mL)	<10,0	$\leq 10,0$

CALIBRATOR

1	9	9	0	2	5	6	8	1	2	2	2	5	6	3	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *[Signature]*



CONTROL SERUM

1	9	9	0	2	5	6	1	3	2	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *[Signature]*



This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. *[Signature]*

DATE 13.06.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.