

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554



QUALITY CONTROL CERTIFICATE

KIT NAME CALPROTECTIN - CHORUS

Code	86600/C03		
Lot No.	033	2025-05	2024-05

REAGENTS

Label	Description	Lot.	Exp.	NO
D9500	CALPROTECTIN	033	2025-05	2X6
D9510	CALPROTECTIN	033	2025-05	1X4
B9510	CALIBRATORE CALPROTECTIN	033	2025-10	1
C9510	CONTROL CALPROTECTIN	033	2025-10	1
G9502	SAMPLE DILUENT	033	2026-05	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	360	≥ 100
Control (mg/kg)	144,8	90,0-250,0

CALIBRATOR

9	0	0	0	3	3	6	2	2	0	2	4	3	9	3	3	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *Adriano*



CONTROL SERUM

9	0	0	0	3	3	6	1	8	3	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *Adriano*



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *Adriano*

DATE *17-06-2024*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.