

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554



QUALITY CONTROL CERTIFICATE

KIT NAME MYCOPLASMA PNEUMONIAE IgG - CHORUS

Code 81034

Lot No. 027 2025-04 2024-04

REAGENTS

Label	Description	Lot	Exp.	NO
D2252	MYCOPLASMA PNEUMONIAE IgG	027	2025-04	6x6
B2253	CALIBRATORE MYCOPLASMA PNEUMONIAE IgG	027	2025-04	1
C2252	CONTROL SERUM MYCOPLASMA PNEUMONIAE IgG	027	2025-04	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	410	≥ 100
Control Serum (UA/mL)	26,0	20,0-43,3
Negative Control (UA/mL)	<10	<10

CALIBRATOR

0 6 6 0 2 7 6 7 1 6 2 9 4 4 4 7 0 0

Label checked with barcode reader YES Operator Signature MB

066027671629444700
MYCOPLASMA PNEUMONIAE IgG
CALIBRATOR 0.425 mL IVD
LOT 027
 2025-04 2024-04
DIESSE Monteriggioni Italy 0123

CONTROL SERUM

0 6 6 0 2 7 6 1 6 3 0 0 0 0 0 0 0 0 1

Label checked with barcode reader YES Operator Signature MB

066027616300000001
MYCOPLASMA PNEUMONIAE IgG
CONTROL+ 0.425 mL IVD
LOT 027
 2025-04 2024-04
DIESSE Monteriggioni Italy 0123

This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents [Signature]

DATE 17.06.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.