

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554



QUALITY CONTROL CERTIFICATE

KIT NAME RESPIRATORY SYNCYTIAL VIRUS IgG - CHORUS

Code	81036		
Lot No.	042	2025-06	2024-06

REAGENTS

Label	Description	Lot.	Exp.	NO
D7661	Respiratory Syncytial Virus IgG	042	2025-06	6X6
B7661	CALIBRATORE Respiratory Syncytial Virus IgG	042	2025-06	1
C7661	CONTROL SERUM Respiratory Syncytial Virus IgG	042	2025-06	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	486	≥ 100
Control Serum (Ndx)	1,5	1,4-2,6
Negative Control (Ndx)	0,0	$\leq 0,7$

CALIBRATOR

1	5	9	0	4	2	6	0	0	0	0	0	5	0	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature MB



CONTROL SERUM

1	5	9	0	4	2	6	1	4	2	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature MB



This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.

DATE 20/06/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.