

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554



QUALITY CONTROL CERTIFICATE

KIT NAME INFLUENZA B IgG -CHORUS

Code	81193		
Lot No.	042	2025-06	2024-06

REAGENTS

Label	Description	Lot.	Exp.	NO
D7532	INFLUENZA B IgG	042	2025-06	6X6
B7532	CALIBRATORE INFLUENZA B IgG	042	2025-06	1
C7532	CONTROL SERUM INFLUENZA B IgG	042	2025-06	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	330	≥ 100
Control Serum (Ndx)	3,8	2,4-4,5
Negative Control (Ndx)	0,0	$\leq 0,7$

CALIBRATOR

0	7	2	0	4	2	6	0	0	0	0	0	2	9	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature HB



CONTROL SERUM

0	7	2	0	4	2	6	2	4	4	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature HB



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents Abun

DATE 20/06/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.