

DIESSSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



QUALITY CONTROL CERTIFICATE

KIT NAME **tTg-A - CHORUS**

Code	86058		
Lot No.	026	2025-04	2024-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D3431	tTg-A	026	2025-04	6X6
B3431	CALIBRATORE tTg-A	026	2025-04	1
C3431	CONTROL SERUM tTg-A	026	2025-04	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	553	≥ 100
Control Serum (AU/mL)	33,6	25,5-51,2

CALIBRATOR

3	3	9	0	2	6	6	2	2	0	3	7	4	5	4	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature

339026622037454900
tTg-A
CALIBRATOR 0.175 mL IVD
LOT 026
2025-04
DIESSSE Monteriggioni Italy CE

CONTROL SERUM

3	3	9	0	2	6	6	1	7	3	2	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature

339026617320000001
tTg-A
CONTROL 0.425 mL IVD
LOT 026
2025-04
DIESSSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 21.06.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.